



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	MEASUREMATIC INDIA, 9,14TH CROSS STREET, MAHARAJA NAGAR, TIRUNELVELI, TAMIL NADU, INDIA	Page No	1 of 4
Accreditation Standard	ISO/IEC 17025:2017	Last Amended on	02/11/2023
Certificate Number	CC-2211		
Validity	14/05/2023 to 13/05/2025		

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current @50Hz	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	10 mA to 100 A	0.025%
2	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Energy (Reactive) 3phase 3Wire @ 40V to 300V, 45Hz to 55Hz,10mA to 10 A, Lead/Lag (0.5 to 1)	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	0.2 VARh to 5.2 kVARh	0.05%
3	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Energy (Active) 1 Phase /3 phase 4 Wire @40V to 300V, 45Hz to 55Hz ,10mA to 100 A, PF Lead/Lag (0.5 to 1)	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	0.2 Wh to 90 kWh	0.05%
4	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Energy (Active) 1 Phase /3 phase 4 Wire @40V to 300V, 45Hz to 55Hz ,10mA to 100 A, PF Lead/Lag (0.5 to 1)	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method:	0.2 VARh to 90 kVARh	0.05%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	MEASUREMATIC INDIA, 9,14TH CROSS STREET, MAHARAJA NAGAR, TIRUNELVELI, TAMIL NADU, INDIA	Page No	2 of 4
Accreditation Standard	ISO/IEC 17025:2017	Last Amended on	02/11/2023
Certificate Number	CC-2211		
Validity	14/05/2023 to 13/05/2025		

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Energy (Active) 3phase, 3Wire @ 40V to 300V, 45Hz to 55Hz,10mA to 10 A, Lead/Lag (0.5 to 1)	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	0.2 Wh to 5.2 kWh	0.05%
6	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage @50Hz	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	40 V to 300 V	0.025 %
7	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Frequency	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	45 Hz to 55 Hz	0.030 Hz
8	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Power Factor @ 40V to 300V ,10mA to 100 A, 50 Hz	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	0.1 PF to 1 PF (Lead & Lag)	0.025 PF



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : MEASUREMATIC INDIA, 9,14TH CROSS STREET, MAHARAJA NAGAR, TIRUNELVELI, TAMIL NADU, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2211 **Page No** 3 of 4

Validity 14/05/2023 to 13/05/2025 **Last Amended on** 02/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Current @50Hz	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	10 mA to 100A	0.034%
2	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Energy (Active) 1 Phase /3 phase 4 Wire @40V to 300V, 45Hz to 55Hz ,10mA to 100 A, PF Lead/Lag (0.5 to 1)	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method:	0.2 Wh to 90 kWh	0.05 %
3	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Energy (Active) 3phase 3Wire @ 40V to 300V, 45Hz to 55Hz,10mA to 10 A, PF Lead/Lag (0.5 to 1)	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	0.2 Wh to 5.2 kWh	0.05 %
4	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Energy (Reactive) 1 Phase /3 phase 4 Wire @40V to 300V, 45Hz to 55Hz ,10mA to 100 A, PF Lead/Lag (0.5 to 1)	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	0.2 VARh to 90 kWh	0.05 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : MEASUREMATIC INDIA, 9,14TH CROSS STREET, MAHARAJA NAGAR, TIRUNELVELI, TAMIL NADU, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2211 **Page No** 4 of 4

Validity 14/05/2023 to 13/05/2025 **Last Amended on** 02/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Energy (Reactive) 3phase 3Wire @ 40V to 300V, 45Hz to 55Hz,10mA to 10 A, PF Lead/Lag (0.5 to 1)	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method:	0.2 VARh to 5.2 kVarh	0.05 %
6	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage@50Hz	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	40 V to 300 V	0.028 %
7	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Frequency	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	45 Hz to 55 Hz	0.030 Hz
8	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Power Factor @ 40V to 300V ,10mA to 100 A, 50Hz	Using Three Phase Reference Standard with Three Phase Power Source by Comparison Method	0.1 PF to 1 PF (Lead & Lag)	0.03 PF

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.